



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of 5(Five) Different medical equipment at different Hybrid CCUs  
across the state of West Bengal  
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-593/2023

Dated-03.10.2023

### Amendment-I

## REVISED TECHNICAL SPECIFICATION

### Schedule-V

#### Nebulizer

Specification	Specification Name	Bid Requirement (Allowed Values)
GENERAL	<b>Weight (kg)</b>	Less than 1 Kg
FUNCTIONAL	<b>Air Flow in Liter Per Minute (LPM)</b>	5
	<b>Noise level (db)</b>	0.0 - 55.0 Or lower
	<b>Size of uniform Particle (MMAD), (micron)</b>	0.5 - 5.0 Or lower
	<b>Technology</b>	Aerosol with piston electro compressor type
	<b>Nabulizing Rate</b>	0.2 to 0.5 ml/minute for 10 LMP
CONSTRUCTION	<b>Air tube and Air Filter and kit with mask(child and adult), (minimum 5 nos of air filters along with each equipment)</b>	Yes
ELECTRICAL	<b>Rechargeable battery 12 Volt/150AH</b>	No

<b>Specification</b>	<b>Specification Name</b>	<b>Bid Requirement (Allowed Values)</b>
	<b>Lead Acid SMF Type</b>	
	<b>Moulded Plug power cord length, ISI Marked (Min)</b>	2.0
	<b>4 Hour Back Up of Battery</b>	Yes/No
Certificate	European CE(4 digit notified body)/CE-IVD/US FDA/BIS/ CDSCO approved	