



## Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited**  
**Swasthya Sathi**  
**GN-29, Salt Lake, Sector-V**  
**Kolkata-700091**

Phone No (033) 40340308/320  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

**Procurement of medical equipment for planned 79 nos. Hybrid CCUs of the Government of West Bengal**

(Submission of Bid through *online*)

2<sup>nd</sup> call of the bid reference no. **WBMSCL/NIT-415/2021 (Schedule-I), Dated- 09.12.2021**

**Bid Reference No.: WBMSCL/NIT-65/2022**

**Dated-23.02.2022**

### Schedule – I

## AMENDMENT-I

### FEEDING PUMP

- Flow Rate Range: 100-550 mL/hr (increments of both 1mL/hr and 5 mL/hr). Upto 100mL minimum 1mL/hr
- Feeding Mode: Continuous/**Intermittent**
- Display: LCD
- Alarm: Technical error, set volume reached, downstream occlusion, upstream occlusion, bag empty, battery empty
- Extra Features:- Priming - 3 Modes: Automatic, Semi-Automatic, Manual
  - Set Volume Activation and Deactivation
  - Alarm for Set Volume Reached
  - Data Recorder/Log Details of :- Total Feed Given,  
Flow Rates of Feed  
Time & Date of Last Feed
  - Flow Rate Accuracy :- 7% at 50mL/hr
  - Low Noise While Pumping
  - Auto Anti Free Flow System

- Sensor Technology for Faster Occlusion
- Feeding Volume:- 100mL @ 1 mL increments  
100mL-5000mL @ 5mL increments
- Easy to Install
- Empty Bag/Air in Line Detection Time at 50mL/hr < 5 min With medical water
- Battery: NiMH , Battery level indication to be displayed, charging time upto 7 hrs:- runs for 24 hrs @ 100mL/hr 5% at 130 mL/hr
- Electrical: 220-240V/50 Hz with Indian Plug
- Compliance: USFDA / CE 4 digit Notarised body 93/42 EEC Medical Directive, EN/IEC 60601-1: 2006 & 60601-1-2:2007
- Availability of flushing facility
- Warranty: 2 years