



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Procurement of equipment /item for the Sub-Centre(SC), Primary Health Centre(PHC), Block Public Health Unit (BPHU) and Urban Primary Health Unit (UPHC) of the State of West Bengal under 15th Finance Commission.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-149/2023

Dated -31.03.2023

AMENDMENT-II **REVISED TECHNICAL SPECIFICATIONS**

Sl. No.	Item	Technical Specification
1	Artery Forceps (Straight)	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided 2. Jaw Pattern: Toothed 3. Overall Length Minimum 6 inches
2	Gauge Cutting Scissor	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided 2. Jaw Pattern: Straight 3. Overall Length 10 inches
3	Stitch cutting scissor	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided 2. Size: 6 inch 3. Jaw Pattern: Curved 4. Lightweight 5. Flawless finish
4	Needle Holder	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided

Sl. No.	Item	Technical Specification
		2. Jaw Pattern: Cross Serated 3. Size: 6 inches
5	Suture Needle Curved	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided
6	Suture Needle Straight-10	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided
7	Foreign body hook (ENT)	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided
8	Vaginal Speculum (Set)	1. Material Grade Stainless Steel 2. Length (mm) 175 3. Size: Medium 4. Speculum should be free from rough edges, pits, cracks, burrs, defects 5. All edges should be smooth and even 6. Speculum should be Autoclavable
9	Mouth Gag	1. Material Grade: Stainless Steel Surgical grade (SS 410) Necessary certificate to provided 2. Overall length 14 cm
10	Dental mirror with probe	Type: Mirror With Handle Shape of Mirror: Round Diameter of Mirror (mm): 18 or more Material of housing of the mirror: Medical Grade Stainless Steel Mirror should be With minimum 10 Cm Length of Handle Mirror should be detachable from the handle Thick Non-Slip Handle for perfect grip Autoclavable
11	Tooth Forceps (Dental)	1. Dental Forceps for extraction of tooth 2. Surgical Grade Stainless steel (SS 410) Necessary certificate to provided 2. Autoclavable 3. Minimum Handle length 10 cm
12	Tuning fork	1. Frequency 256 Hz 2. Made up to good quality rust free stainless steel. 3. Should have minimum Decay time of 40 sec. 4. Footplate should be present
13	Ambu Bag (Paediatric Size)	1. Silicon Resuscitator Infant self inflating silicone bag with mounts incorporating Reservoir valve and side feed oxygen inlet (200 ml – 650 ml). 2. Bag, reservoir and mask should be Autoclavable 3. Size 0, 1 facemask with silicone cuff. 4. Oxygen tube & 1000 ml Reservoir bag 5. Pressure Limiting Valve adjustments for

Sl. No.	Item	Technical Specification
		performance 6. Carry bag.
14	Nebulizer	1. Light Weight ,Portable and Compact : Yes 2. Air Flow in Liter Per Minute (LPM) : 10 3. Nabalizing Rate : 0.2 ml/minute for 5 LMP 4. Air tube and Air Filter and kit with mask
15	Measuring Tape (MUAC)	1. 3 colour Mid-Upper Arm Circumference(MUAC) tape for accurate assessment of nutritional status for Children in WHO standard 2. Size of tape (LxW)(mmxmm) : 335x10 3. MUAC Tape Weight: 5 gms 4. Measures length in centimeters: Up to 33 5. Graduation of measuring scale in Centimeters: Every 1 mm 6. Marking colour 0 to 11 decimal 5 Cm: Red 7. Marking colour 11 decimal 5 to 12 decimal 5 Cm: Yellow 8. Marking colour from 12 decimal 5Cm onwards: Green 9. Non-stretchable PVC material, Non-tearable, Flexible and foldable, Washable with water, Simple for measurement 10. Calibration certificate from concerned weights and measurement controllerate to be provided
16	Stadiometer	1. Metallic Height Measuring scale with movable upper part which can be fixed according to the height of the child measuring range 0 - 220 cm; Graduations 1 MM. 2. Firm flat vertical surface 3. Fold-up mechanism, easy to transport, compact design for heavy duty use 4. Double sided graduation for easy measurements 5. Board material: Stainless Steel 6. All the items/ models shall be duly approved by concerned weights and measurement authority & the units offered for the supplies shall be duly calibrated as per the procedure of department of legal metrology