

### **Notice Inviting e-Tender**

# West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

# Procurement of Medical Items for the Blood Centres at Hospitals and Medical Colleges of Govt. of West Bengal

(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-193/2024 Dated-11.03.2024

#### **Amendment-II**

# REVISED TECHNICAL SPECIFICATION ITEM-12

## **PQS DIGITAL TEMPERATURE MONITOR**

- 1. Temperature Monitor Certified for Freezers, Refrigerators, Ambient / Room Incubators, Water Baths and Block Heaters.
- 2. Each thermometer is sterilized and certified, traceable to NIST.
- All sensor probes are housed in an unbreakable plastic bottle filled with a Bio-Safe buffering media which maintains readings from sudden temperature changes.
- 4. An audible alarm will sound when the temperature rises above or goes below the set temperatures.
- 5. The thermometer and the bottle probe have Dual Magnets allowing attachment to doors or walls.
- 6. Other features include a flip-open stand.
- 7. The thermometer is <sup>0</sup>C and <sup>0</sup>F switchable.
- 8. Range:  $-12^{\circ}$ C to  $+10^{\circ}$ C.
- 9. Warranty: 2 years;
- 10. Product Certification: European CE / US FDA / BIS certified

#### \*\*Note Applicable for item no. 1 to 22:

- 1. Installation Quality (IQ), Operational Quality (OQ), Performance Quality (PQ) certificate to be submitted during installation of the machine.
- 2. 3<sup>rd</sup> party calibration certification should be also be provided during installation