



Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091**

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

**Supply and Commissioning of Medical equipments for setting up of Sports Medicine unit
at Govt. Medical College & Hospital of the Govt. of West Bengal**
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-36/2022

Dated-28.01.2022

The following amendments have been made in the technical specifications of the tender document,

Amendment – V

(In Page: 37)

Schedule - II

ANTI GRAVITY TREADMILL

1. The unit should be useful for rehabilitation of lower limb extremities (hip, knee, ankle or foot) and for paediatrics as well as adult patients.
2. The unit should have work on the principal of Differential Air Pressure (DAP) technology and have LCD/LED touch screen display.
3. The unit should have forward speed of the treadmill from 0-15 kmph and reverse speed of the treadmill 0-3kmph.
4. The unit should have speed adjustment increment of 0.16 kmph.
5. The unit should have inclination of the treadmill 0-15%.

6. The unit should have Air Pressure mechanism for partial weight bearing system.
7. The unit should have Body weight reduction should be 0 -80% with 1% increment.
8. The unit should have User weight capacity (for small, medium & large) upto 180kg.
9. Paediatric Height: 152-190 cm
10. Medium user height: 168-200 cm
11. Large user height: 185-220 cm
12. User hip size (for small, medium & large): Width min. 47 cm, circumference min. 147 cm
13. The unit should be USA FDA or European CE approved (4 digit).
14. The unit should work on 220V/50Hz
15. The system should be supplied with CVT/Servo Voltage Stabilizer of required rating.