



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

**Procurement of Medical Items for the Blood Centres at Hospitals and Medical
Colleges of Govt. of West Bengal**
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-193/2024

Dated-11.03.2024

Amendment-VII

REVISED TECHNICAL SPECIFICATION

ITEM-17

ELECTRONIC WEIGHING SCALE

Feature	Specification
1. Input supply	100 to 240V AC, 0.7A, 50 to 60 Hz
2. Power consumption	5W (Max)
3. PAN	Single PAN
4. Ingress of water protection rating	IPX1 protection against dripping water
5. Transducer	Load cell
6. Weighing range	Up to 5 kg
7. Accuracy	1g or 1mL
8. Interface ability	Integrated RS232 connection port
9. Tear function	Automatic initialization/manual
10. Display	LCD (backlit)
11. Switch	Power ON/OFF

12.Keyboard	Weight /Volume Conversation; volume setting(up/down); Component Selection (up/down); tare function balancing of bags; enter/escape
13.Indications	Tare function: LED g/mL: LED Over range: LCD Set weight/volume achieved: LCD
14.Product certification	CE / USFDA / BIS certified