



## Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited  
Swasthya Sathi  
GN-29, Salt Lake, Sector-V  
Kolkata-700091

Phone No (033) 40340307/319

E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in) / [to@wbmsc.gov.in](mailto:to@wbmsc.gov.in)

Procurement of equipment /item for the Sub-Centre(SC), Primary Health Centre(PHC), Block Public Health Unit (BPHU) and Urban Primary Health Unit (UPHC) of the State of West Bengal under 15th Finance Commission.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 144/2022

Dated-12.04.2022

The following amendment have been made in the tender document,

### Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

#### Hot Air Oven

- i) Outside made of mild steel with powder coated paint.
- ii) Inside made of stainless steel Grade 304.
- iii) The double walled door with asbestos gasket lining is fitted on heavy brass cast chromium plated hinges and has a chromium plated latch type lock.
- iv) Temperature can be controlled from few degrees above ambient to 250°C by thermostat.
- v) Heating elements are placed in the ribs at the bottom and sides.
- vi) Built in L-Shape prismatic / digital thermometer, with two shelves, works on 220/230V AC supply with air circulating fan.

- vii) Option: Electronic Digital Temperature controller cum indicator, inner chamber size 14X14X14 inch / equivalent volume.
- viii) Safety Certification: CE/ US FDA/ BIS
- ix) Warranty 3 years