



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and commissioning of various ophthalmic equipment across different Health Facilities of the Govt of West Bengal under the project NPCB and VI

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-293/2025

Dated-16.04.2025

Amendment I

Revised Technical Specification

Schedule No.V Streak Retinoscope

1. The Streak retinoscope should be Parastop/ cylinder and the good fundus reflex and easier detection of the neutralization point.
2. It should be homogeneous xenon/halogen illumination for precise streak image quality offers an easy and quick observation of the fundus reflex.
3. It should be precise easy selection of a parallel beam.
4. It should be stepless dimming from 3% to 100% or better with practical one finger operation.
5. It should have all rechargeable metal handle and the voltage range should be 3.5V or more.
6. It should have ergonomic shape and protect the orbita of examiner from stray lights.
7. It should have metal control which should be long lasting.
8. It should have single control for vergence and rotation for comfortable operation.
9. It should be maintenance free and dustproof.

10. It should have integrated or similar polarization filter eliminates stray light and internal reflexes for a brighter fundus reflex.

Safety & Standard: Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable