



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical equipment for Mother and Child Hub at Canning Sub
Divisional Hospital, South 24 Pgs
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-331/2022

Dated-25.07.2022

The following amendment has been made in the bid document:

Amendment – IV (Page No.- 35)

Item – 4

Ventouse

1. Ventouse should have Foot operated electric Sucker.
2. Pressure 600 mm Hg with user control flow change buttons and selectable airflow rate of 40-60 liter/min.
3. Power consumption 40 watt - Pump noise within 40-60 db.
4. **Cup (Silicon Rubber):**
Silastic cup - 50 mm and 60 mm one piece each.
Smooth external surface i.e., smooth longitudinal ridge to observe rotation.
Suitable for autoclave sterilization.
5. Accurate functioning
Low maintenance
Tough construction
6. Ventouse should have Digital / Analog display of pressure system.
7. Ventouse should have all silicon tubes
8. Ventouse should have Trolley
9. Should be European CE / USFDA / UKCA / **UK Certification** / BIS certified.