



## **Notice Inviting e-Tender**

**West Bengal Medical Services Corporation Limited**  
**Swasthya Sathi**  
**GN-29, Salt Lake, Sector-V**  
**Kolkata-700091**

Phone No (033) 40340307/320  
E mail: [procurement@wbmsc.gov.in](mailto:procurement@wbmsc.gov.in)

**Supply and Commissioning of different types of Protective Instruments and Gears for the purpose of AERB Compliance across all the Health Facilities under Govt. of West Bengal**  
(Submission of Bid through *online*)

**Bid Reference No.: WBMSCL/NIT-137/2025**

**Dated-18.02.2025**

**2<sup>nd</sup> call to Bid Reference Number WBMSCL/NIT-035/2025 Date-10.01.2025**

## **Amendment-I**

# **REVISED TECHNICAL SPECIFICATION**

**Page 48**

### **Item No. – 5** **Movable protective Shield**

- a) It should be mobile and sturdy.
- b) Movable Lead Shield should have Lead equivalence of 1.5 mm (Pb).
- c) Adjustable Movable Lead Shield can be set between 36 inches and 72 inches high.
- d) Dimensions of the Lead Shield should be 50 x 50 (cms)
- e) It should be convenient to move in any direction.
- f) CDSCO