



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of different medical equipment for Department of Neuro OT and ICU at Bangur Institute of Neurosciences, Annex-I, IPGME&R SSKM Hospital

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-069/2024

Dated-19.01.2024

AMENDMENT-I

Schedule-I

REVISED TECHNICAL SPECIFICATION

Mobile C-arm with fluoroscopy Machine

The System should have below mentioned Specification:

1. I.I.T.V System

- a) The image intensifier should be of laser series.
- b) It should be of 9 inches filed i.e. 9 inches / 6 inches / 4.5 inches.
- c) The centre resolution should be minimum 48 lp/cm.
- d) The Circular grid should be fixed on the image intensifier to improve image quality

Or, Flat panel detector

- a) Receptor type should be of Amorphous Silicon technology
- b) Conversation screen should of Csl
- c) FPD with 21x21 cm size or more should be provided
- d) Image matrix should be 1k x 1k or more
- e) Pixel pitch should be 205µm or less

f) ADC conversation should be 16 bit or more.

2. **CCD / CMOS Camera:**

- a) It should be ruggedly built and should be of good design
- b) It should have 1 or 2 separate steering for controlling back and front wheel movements
- c) It should also have the below mentioned movements.
 - Horizontal travel should be minimum 200mm
 - Orbital movement should be above 125°
 - Panning movement should be $\pm 12.5^{\circ}$
 - Vertical movement should motorized of 400 mm
 - Focus to distance should be 900mm
 - C-Arm rotation should be $\pm 180^{\circ}$ preferably $\pm 360^{\circ}$

3. **C-ARM Stand:**

- The CCD / CMOS camera should be $\frac{1}{2}$ inch
- It should have resolution of 1k x 1k minimum

4. **Monitors:**

- a) Medical grade monitor minimum 19 inches more on trolley – 2 Nos.
- b) The monitor trolley should be provided for mounting 2 monitors and should have 2 shelf for keeping memory and stabilizer.

5. **Generator:**

- a) It should be microprocessor controlled digital system with display
- b) It should be of high frequency with output of minimum 3.5 KW and frequency of 40 Khz. (Preferably 100KHz-200KHz)
- c) The KV should be from 40 to 110 KV.
- d) The fluoroscopic mA should be from 0.3 to 3.0 mA or wider.
- e) The system should have fluoroscopy mode like
 - Manual Fluoro mode and Continues Fluoro mode.
 - Pulsed fluoro mode with facility to select time interval between the pulse from 1 pps to 10 pps or more
 - Auto Dose Rate Control in fluoroscopy mode by which mA & KV should be set automatically as per the thickness of the organ
 - Manual KV slection during fluoroscopy also should be available
 - Boost fluoroscopy mode optional/High definition fluoroscopy
- f) The digital fluoroscopic timer should be incorporated with arrangement of auto cut off of exposure after 300 secs.
- g) The radiographic mAs ranges should be from 20 to 30 mAs or more

h) The X-ray tube should be dual focus stationary anode. The focal spot of the tube should be

1. Small: 0.6mm x 0.6mm
2. Large: 1.8mm x 1.8mm or less

It should have mono block / tube housing heat storage capacity of 200 KHU or more. It should also have inherent filtration of 0.5mm or more A1 eq.

- i) The system should have backlit LCD display of fluoro mA, KV, timer & radiography mAs should be provided
- j) The reversal, image rotation, functions should be operatable either from control panel or with a remote control
- k) Memory functions like store recall/image transfer should be operatable from control panel as well as from memory unit.
- l) There should be independent selection of mA and KV & mAs.
- m) The control should have indicator for power, Overload, X-ray & tube heating.
- n) The system should be upgradable to latest functions

6. Image Memory

- a) Digital image Processing & Memory system with PC or a USB Drive.
- b) The system should have DVD recording facility as externally or internally
- c) It should have 50000 images or above
- d) It should have at least 100 permanent images storage capacity
- e) It should have image integration function to reduce the image noise
- f) Should be capable to copying images to Pen drive.

Essential Accessories

- a) Lead aprons (thickness 0.5mm), Thyroid Shield, Lead Goggles (10 nos each)
- b) Lead apron stand - 02 Nos. & for each stand 5 nos. Hanger should be supplied
- c) Servo stabilizer - 1

7. Should be AERB approved

8. The system should be DICOM compatible.

The product should have US FDA / European CE (4 digit notified body) /BIS