



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical equipment for 02 (two) new Government Medical College & Hospital, Dept. Of Pathology
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-247/2025

Dated-28.03.2025

2nd call of bid reference no. WBMSCL/NIT-198/2025 Dated-24.04.2023

Amendment-II

REVISED TECHNICAL SPECIFICATION

Item No.-02

PARAFFIN EMBEDDING BATH

1. Paraffin keeping bath capacity: 5000mL. Paraffin flowing tube temperature: 50~70°C
2. Paraffin dropping board temperature: 50~70°C. Tools bath temperature: 50~70°C
3. Tissue keeping tank temperature: 50~70°C. Tissue cassettes keeping bath temperature: 50~70°C
4. Pincer hole numbers: 4 stations. Tissue cassettes keeping bath size: 260×200×50mm
5. **Voltage and frequency: AC220V±22V 50Hz.**
6. LED displays current temperature, preset value, start/end of time and so on.. Large working surface can put 100 cassettes. Attaining working temperature after switching on 10 minutes.
7. Certifications: Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable