



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
E mail: procurement@wbmsc.gov.in

PROCUREMENT OF OPERATING HYSTEROSCOPE WITH ACCESSORIES & ENDOMAT AT
WALSH SDH SREERAMPORE

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL /NIT-219/2020

Dated –11.11.2020

(2nd call of Bid Reference No.: WBMSCL/NIT-202/2020, Dated- 13.10.2020)

The following amendments have been made in the tender document,

Amendment – II (Page No. 34 to 35)

Sl. No.	Description	Qty	Covered under warranty & CMC
1	Hysteroscopy Diagnostic Sheath for 2.7 mm to 2.9 mm telescope	01 Nos.	Y (2 years)
2	Full HD Telescope 2.7 mm to 2.9 mm 30 Degree with Rod Lenses, Optical system, Autoclavable with incorporated fiber optic light transmission length minimum 200 mm; Should have European CE/ US FDA certified	01 Nos.	Manufacturing warranty for 1 year
3	Operative sheath (Inner & Outer) for 2.7 mm to 2.9 mm forward oblique telescope 30 degree with 5 Fr instrument Channel Auto lock system with Discreet inflow & outflow channels	01 Nos.	Y (2 years)
4	Semi Rigid Scissor Single Action 5Fr	01 Nos.	Y (2 years)
5	Semi Rigid Alligator / Grasping Forcep 5 Fr	01 Nos.	Y (2 years)
6	Semi Rigid Biopsy Forcep 5 Fr	01 Nos.	Y (2 years)
7	Bipolar L Hook 5 Fr	01 Nos.	Y (2 years)
8	Continuous Flow Inner outer Sheath with Obturator For 2.7 mm to 2.9 mm Telescope for TCRE	01 Nos.	Y (2 years)
9	Mono as well as Bipolar working Element	01 Nos.	Y (2 years)

Sl. No.	Description	Qty	Covered under warranty & CMC
10	Cord For Working Element	01 Nos.	Y (2 years)
11	Bipolar Cutting Loop	01 Nos.	Y (2 years)
12	Bipolar Ball Electrode	01 Nos.	Y (2 years)
13	Collins knife	01 Nos.	Y (2 years)
14	Tubal Recanalization Guide wire	01 Nos.	Y (2 years)
15	Ellik Evacuator	01 Nos.	Y (2 years)
16	Full HD Telescope 4 mm 30 Degree with Rod Lenses, Optical system, Autoclavable with incorporated fiber optic light transmission length minimum 200mm; Should have European CE	01 Nos.	Manufacturing warranty for 1 year
17	Diagnostic Sheaths	01 Nos.	Y (2 years)
18	Operative sheath (Inner & Outer) for 4 mm forward oblique telescope 30 degree with 5Fr to 7 Fr instrument Channel Auto lock system with Discreet inflow & outflow channels Operative Sheath Continuous flow	01 Nos.	Y (2 years)
19	Mono polar Bug Bee Electrode	01 Nos.	Y (2 years)
20	Semi Rigid Scissor Single Action 5Fr to 7 Fr	01 Nos.	Y (2 years)
21	Semi Rigid Alligator/Grasper 5Fr to 7 Fr	01 Nos.	Y (2 years)
22	Semi Rigid Biopsy Forcep 5Fr to 7 Fr	01 Nos.	Y (2 years)
23	Continuous Flow Inner Outer Sheath For 4 mm Telescope for TCRE	01 Nos.	Y (2 years)
24	Mono as well as bipolar Working Element. For 4 mm Telescope	01 Nos.	Y (2 years)
25	Loop Straight, 90 Degree	01 Nos.	Y (2 years)
26	Collins knife	01 Nos.	Y (2 years)
27	Endomat/ Hysteropump <ul style="list-style-type: none"> ➤ Flow & Pressure controlled Irrigation System ➤ Liquid Pressure Monitoring Technology ➤ Disposable Pressure transducer to avoid Patient Fluid Contamination ➤ Real Time Pressure & Flow Monitoring of Intra-Uterine Cavity I specific surgical site ➤ Built-in Low Noise irrigation pump ➤ Surgery Specific Custom Modes : Hysteroscopy & versatile ➤ Preferably indicator for amount of consumed fluid by the patient ➤ Preferably for Saline & Glycine Mode settings Alarm indication for consumed fluid after safety threshold ➤ Priming / Flush for removing air bubble from tubing at initialization ➤ Touch Screen LCD Interface <p> Pressure Range: 100 mmHg to 400 mmHg \pm 10% Flow Range: Irrigation : 100mL to 2000mL \pm 10% Pressure Setting: Irrigation : 10mmHg 1 step Flow Setting: Irrigation : 100mL 1 step Operating Modes: Only Irrigation Mode Filter: Disposable Cartridge Filter (0.2 micron) Tubing : Silicon Tubing Tube Specifications: Silicon tubing for Irrigation and Suction of 3Mtr Touch Control Panel: 5" LCD Screen with brightness control To be supplied with reusable Irrigation Tubing Set </p>	01 Nos.	Y (2 years)

Amendment – III

(Page No. 3)

5 (a) Time for Supplies & Commissioning of work from the date of issuance of Award of Contract

Item	Time
Operating Hysteroscope with accessories and Endomat	45 Days