



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340308/319
 E mail: procurement@wbmsc.gov.in

**PROCUREMENT OF DIFFERENT MEDICAL EQUIPMENTS FOR MINI MCH HUB AT SUJAPUR
 PHC & SILAMPUR RH OF MALDA DISTRICT**

Bid Reference No.: WBMSCL /NIT-97/2022

Dated –11.03.2022

(2nd call to Bid Number: GEM/2021/B/1034925, Dated: 18-02-2021, GEM/2021/B/1034846, Dated: 18-02-2021 & WBMSCL /NIT-24/2021, Dated –21.01.2021)

The following amendment has been made in the tender document,

SCHEDULE-III

Amendment –I

(Page No. 38)

| <u>Technical Specification of Breast Milk Expressor</u> | | |
|---|--|--|
| Specification | Specification Name | Bid Requirement (Allowed Values) |
| Standards | Conformity to Safety Standards | IEC 60601-1 or BIS equivalent |
| Performance Parameters | Pump Configuration | Electrical |
| | User | Multiple user |
| | Suction control range (mm Hg) | 30mmHg to250mmHg and adjustable, 50mm Hg to250mm Hg and adjustable |
| | Pumping frequency (cpm) | Pumping frequency 30 to 80 Cpm and user adjustable., Pumping frequency 40 to 130 cpm |
| | Able to express milk from both breasts simultaneously | Yes |
| | Cushion inserted inside the breast cup so that it does | No, incase of silicon soft material provided |

| | | |
|---|---|---|
| | not hurt the mother | |
| | Display | Digital Display |
| | Type of Display | LCD |
| | Parameters displayed | Regulated Speed,Suction and timer |
| | Operational Messages for operational indications | Yes |
| | Pumpset properly attached | Yes |
| | Unit is in service | Yes |
| | Double alternating pumps/double cycling pumps | Yes |
| | Automatic shut-off | With / without |
| | Trolley | Yes |
| | Weight of Trolley in kgs | NA (if not provided) |
| | Display shows suction in % of 1-100% | Yes |
| | Soft-touch control panel | Yes |
| | Collection bottles used for storage of milk should be autoclavable and bio compatible | Yes |
| | Dual expression breast bottles set including tubing | Yes |
| | Sets of Reusable collection bottles along-with breast cups | 10 set |
| | Noise level in dB | less than 50 db |
| | Diaphragm | Yes |
| | Number of Diaphragm | 100 no., NA (If not supplied) |
| | Including Valves | Yes |
| | Number of sets of all kinds of tubes | 2 sets |
| Miscellaneous Parameters | Warranty in Years | 3 |
| Additional Specification Parameters - Breast Pump (5 pieces) | | |
| Specification Name | Parameter | Bid Requirement (Allowed Values) |
| Certification: | | European CE or US FDA |