



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical equipment for 02 (two) new Government
Medical College & Hospital, Dept. Of Microbiology

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-486/2024

Dated-30.08.2024

2nd call for items at Sl No. 2 , 5 , 6 and 8 of Bid Reference Number WBMSCL/NIT-155/2023
Date : 31.03.2023

Amendment-I for Item No. 4

REVISED TECHNICAL SPECIFICATION

Item No. 4	Water bath (for Microbiology and serology)	1. Capacity: 10 ltr, 2. Rectangular water bath 3. Digital display, membrane keyboard, temperature range from
-------------------	--	---

		<p>ambient to +5°C, set value memory.</p> <p>4. High temperature consistency (control accuracy 0.5° C), working temperature 90 degree above ambient temperature indicated by flashing LED. Set temperature displayed in addition to actual temperature.</p> <p>5. Fully SS surface with special scratch-proof , leak proof with SS lid ensures highest thermal conductivity.</p> <p>6. Option for thermometer. Inbuilt calibrated thermometer</p> <p>7.Digital Microprocessor</p> <p>Certification: CE / EC / UK Cert / US FDA / BIS</p>
--	--	--