



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited

Swasthya Sathi

GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

**Procurement of Medical Items for the Blood Centres at Hospitals and Medical
Colleges of Govt. of West Bengal**
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-193/2024

Dated-11.03.2024

Amendment-II

REVISED TECHNICAL SPECIFICATION

ITEM-1

INCUBATOR WITH THERMOSTATIC CONTROL (75 LTR.)

- a) Control: Equipped with Microprocessor controlled with a large digital display.
- b) Work Chamber Volume: minimum 75 Litres.
- c) Working temp.: Ambient temp. +5°C to 75°C whereas room temperature is around 20° – 24° C.
- d) Temp. deviation: ±0.5°C.
- e) Display: Digital display located on front panel of the Items.
- f) Easy to clean: A silicon door gasket & removal of shelf supports make thorough cleaning and disinfection of the working chamber easy & safe
- g) Inner casing: Should be made of corrosion resistant 304 grade stainless steel.
- h) Shelves (adjustable): Two perforated shelves made of corrosion resistant 304 grade stainless steel. Standard supply 3 nos. , inter shelves gap – 6” approx.
- i) Heat conduction: **Air circulation.**
- j) Heater on rear / side / bottom of the Items.
- k) Power requirement: 230 V, single phase, 50 Hz.
- l) **Warranty : 2 years**
- m) **Certifications:**

- **Product certification:** CE / US FDA / BIS certified

n) Product should have NABL accredited calibrated thermometer (temp. range 0°-100°C).