



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Flexible Ureteroscope for Dept. of Urology of Medical College & Hospital, Kolkata.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-371/2023

Dated-10.07.2023

2nd call of Bid Reference Number WBMSCL/NIT-305/2023 Dated: 06.06.2023

Amendment-I

TECHNICAL SPECIFICATION

Technical Specification of Flexible Uretero – Renoscope(Reusable)

- It should have working length of 60-75 cm.
- Outer diameter of the shaft should be between 6.0 Fr - 8 Fr.
- It should have Direction of view 0 – 12 degree.
- It should have angle of deflection up to 270 degree **or more** downward and 270 degree **or more** upward is needed.
- It should have a ceramic liner in the distal end of the working channel to protect it from thermal or electrocautery damage.
- It should have integrated shock absorber system.
- It should be waterproof and shaft is immersible in sterilizer solution
- It should adhere to sterilization method with ETO / FO gas / Steris / Sterrad.

- Following compatible accessories should be supplied with this instrument – case for the instrument, pressure compensation cap, and leakage tester and cleaning brush.
- It should be compatible with any existing camera system.
- Grasping Forceps or basket (Renowned company accepted by the Experts)– 4 (1.5-2.5 Fr)
- Biopsy Forceps (Renowned company accepted by the Experts) - 2
- Access Sheath (Renowned company accepted by the Experts):
 - i) 45 cm – (10-12 Fr) / (12-14 Fr) – 6 each (total 12)
 - ii) 35 cm - (10-12 Fr) / (12-14 Fr) – 6 each (total 12)
 - iii) 28 cm – (10-12 Fr) / (12-14 Fr) – 6 each (total 12)
- Warranty : 2 years
- Certificate : European CE (4 digit notified body) and USFDA / BIS