



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Procurement of Ophthalmic equipments for six Medical Colleges and Hospitals

Bid Reference No.: WBMSCL/NIT-468/2022

Dated-28.10.2022

Amendment-I

REVISED TECHNICAL SPECIFICATIONS

Schedule-I
(Page No: 33)

1. Technical Specification of Diode Red (810 nm wave length) Laser Indirect Ophthalmoscope

1. **Weight:** Light Weight (Portable)
2. **Power Requirements:** Should be operable with our standard electrical support.
3. **Cooling:** No external air or water cooling required
Treatment Laser: Semiconductor / **Solid state** diode laser
Wavelength: 810 nm
4. **Delivery Devices and Output Power Ranges:**
 - Trans-scleral Glaucoma Probe: 0-2000 mW
 - Endo Probe Handpiece
 - Laser Indirect **ophthalmoscope**
5. **Operating Microscope Adapter:**

- Spot sizes: Fixed
- 6. **Exposure Duration:** 10 ms to continuous (in graded manner); extended durations with operating microscope adapter.
- 7. **Repeat Interval:** 100 - 700 ms or more and single pulse
- 8. **Aiming laser:**
 - Red semiconductor laser
 - Wavelength: 635 to 650 nm Power: User adjustable 0-<1.0 mW
- 9. **Certifications-**4 digit European CE from notified body/ US FDA .

Amendment-II

In Section I ; Sub Section E(Submission and Opening of Bids) Point No.36 , Sub Point (iii b) Pg 13

“Cost of Consumable items which would likely to be procured in staggered manner or as per requirement of the end user for next 5 years”.-DELETED