

Revised Technical Specifications

LED Examination Light for medical use

Specification	Specification Name	Bid Requirement (Allowed Values)
LIGHT SOURCE DETAILS	Correlated Colour Temperature (CCT)	4000-5000 degree K
	Colour rendering index	80.0 - 100.0 Or higher
	Minimum Light Intensity at working distance	20000, 50000, 100000 Or higher
	Minimum LED Life	50000, 75000, 100000 Or higher
POWER & CONTROLS	Availability of Continuous dimming levels	Yes
	Length of the Power Chord	1.0 Or higher
MOUNTING & ACCESORIES	Mounting options	Mobile Stand
	Wheel locking provision	Yes
	Disposable Sheath	Available
CERTIFICATION	Product complies with IEC 60601-1 and IEC60601-2-41	Yes
Additional Specification Parameters - LED Examination Light for medical use		
Specification Parameter Name		Bid Requirement (Allowed Values)
Warranty		3 years
Certification		CE/ US FDA/ BIS
Warranty		3 years