

Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

Supply and Commissioning of different types of Medical Equipment for the Dept. of General Medicine at Burdwan Medical College & Hospital

(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-538/2025 Dated-23.06.2025

AMENDMENT-I

Schedule-I

REVISED Technical Specification of Echocardiography Machines

- 1. The system should be state of the art with the facility of 2D, M-Mode, CDI, CW Doppler, Power Doppler, Directional power angio, and should have Imaging for abdomen, Cerebrovascular, peripheral vascular & superficial parts imaging, should have mandatory package for Adult & Pediatric Echocardiography.
- 2. System must have more than 7,00,000 / 7 Lac or more digital processing channels.
- 3. System should have dynamic range of 256 dB or more.
- 4. System should be offered with a 21 inch or more High Resolution Flat Panel Medical grade. Display monitor with facility for position adjustments with 8 inch or more touch panel.
- 5. System should have at-least four universal active probe ports with electronic switching facility from key board without probe adapter.
- 6. Operating modes B-mode, M-mode, B/M Mode, Doppler Mode, Color flow, Power Doppler, DCA/DPA, B/Color flow, PW Doppler, CW.
- 7. Probes should be of board band type.
- 8. B Mode & color-flow images should be simultaneously available side by side in real time. Digital zoom facility for region of interest in real time and frozen images.

- Image storage facility on in build hard disc or MOD/CD/DVDRW facility should be available. In built hard disk with minimum capacity of 500 gb or more. System should have extensive image management capability including thumb nail review. Cineloop editing etc
- 10. Cine loop as well as cine scroll facility in B mode with storage of 1000 or more images should be available
- 11. Should have Real Time compound Imaging Technology with Multiple transmitted lines of sight in convex, linear and endocavitary probes. System should have stress echo package.
- 12. Should be offered with speckle Reduction Imaging, panoramic imaging should be available.
- 13. Advanced measurements & calculation package for abdominal, obst./gynae, urology and vascular and intracavitary invention applications should be available
- 14. System should be capable of scanning up to depth of 30cm or more.
- 15. System should be offered with a 2D frame rate if at least 1000 frames/second or more.
- 16. System should have THI & should be able to work in combined mode of harmonic imaging and real time compound imaging to get excellent image quality. The system should offer Tissue Harmonic Imaging in Power Doppler imaging mode for improved sensitivity.
- 17. Automatic real time & frozen tracing of instantaneous peak velocity & instantaneous mean velocity (or frequency) should be available. Triple, Imaging should be standard on the system.
- 18. The system should be DICOM3.0 (or higher version) ready (like send, receive, print, record on CD/DVD, acknowledge etc.) for connectivity to any network, PC/computer etc. In DICOM format. Vendor will connect the machine to existing PACS and to local other laser cameras without additional cost.
- 19. The bidder should submit valid CDSCO Certificate/Registration/License for both the manufacturer(s) and importer(s) as applicable.

System should be offered with the following transducers

- 1. Adult echo Probe of 2 to 5 MHz (+/-1) with Tissue Harmonics Imaging.
- 2. Pediatric Echo Probe of 3 to 8 MHz (+/-1) with Tissue Harmonics Imaging.
- 3. Board band Linear array transducer suitable for Vascular and small parts applications -5 to 12 MHz (+/-1).

Future Upgrade (Optional Prices to be quoted separately):-

- 1. 2D Adult/Paediatric TEE Probe of 2-7 MHz (+/-1).
- 2. Board band Linear array transducer suitable for Vascular and small parts applications -5 to 12 MHz (+/-1).

3rd Party Item to be offered along with system.

- 1. UPS with 30 Min Back Up
- 2. Thermal Printer.