

West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

Supply and Commissioning of ENT Operating Microscope in the Dept. of ENT of different Medical Colleges and Hospital under Govt. of West Bengal

(Submission of Bid through online)

Bid Reference No.: WBMSCL/NIT-512/2025

Dated-18.06.2025

2nd call of Bid Reference Number WBMSCL/NIT-350/2025 Date-30.04.2025

<u>Amendment I</u>

Revised Technical Specification

ENT Surgical Microscope

1. Working distance:

Total variable working distance 200-400 mm or better, continuously variable through multifocal lens through Hand control

2. Continuous Zoom/ Step:

Microscope should have continuous zoom/ step to work at any magnified position with dial indicator to visualize value of magnification factor.

3. Magnification range:

Minimum range up to 17X or better

4. Focusing with LASER attachment:

Fine focusing for lens activated through Hand adjustable

5. Eyepieces:

12.5X wide field eyepiece with dioptric setting +5D to -5D

6. Light Source:

High power Integrated LED illumination min. 5500K illumination.

7. Binocular tube:

Binocular tube for main Surgeon which canbe pushed and pulled offering flexible positing, added magnification and Integrated rotate functionality. Tiltable range is 0⁰-180⁰ or more

8. Hand grip

Ergonomic handle, bilateral, 360°, adjustable with release knob.

9. Lateral Movement

Binocular Turntable for Lateral rotation in ± 12 Degree for neck comfort to the Surgeon

10. Beam splitter:

Integrated Beam splitter for camera

11. Camera:

Integrated Full HD Video Camera that maximum resolution will display & record

- Apochrimatic optics tasted generation for brilliant and true color images and videos
- Highest resolution for current and future high-definition camera systems with APS-C-Format
- Ideally matched to camera and microscope Full format without any annoying vignetting shading
- Compact overall height and minimum weight, no adverse effect on the word with the microscope
- Live full HD-video streaming in 1920 x 1080p via HDMI

12. Display:

32" or more Full HD display system must have remote control operation

13. Recording:

Integrated HD video recording systems with HDD of at least 1 TB with HDMI input/ output. System must have remote control operation.

14. Side Assistant Scope:

Microscope must have stereo co-observation attachment for side assistant for left or right side assistant should also offer with tiltable tube with range of 0 degree to 180 degree or more with 10X or 12.5X Wide field eyepiece with DIOPTERIC Setting +5D to-5D

15. Spot Illumination:

Spot illumination control is required to view small particular anatomy

16. **Asepsis:**

We provide asepsis caps for all controls and handles.

- 17. Lighting with no glare
- 18. Filters to differentiate between tissue & vascularity and tissue & blood.
- 19. Prior to opening of financial bid physical demonstration of their quoted product by the participant bidders at Dept. of ENT of MCH Kolkata is mandatory
- 20. **Standard and Safety :** Valid CDSCO Certificate License for the Manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable