



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320

E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Anaesthesia Workstation for R.G.KAR MCH

(Submission of Bid through *online*)

2nd call of Bid Reference No.: WBMSCL/NIT-498/2022; dated-21.11.2022

Bid Reference No.: WBMSCL/NIT-598/2022

Dated-29.12.2022

Amendment-II

Revised Technical Specification

Anaesthesia Work Station

Parameter / Feature	Threshold Limit / Range
Battery back up in minutes	At least 90 mins.
Gas flow in liter	(0.10 - 12) or more
Vaporiser flow range in liter/minutes	(0.2 -15) or more
Vaporiser chamber capacity in ml.	200 or more

Parameter / Feature	Threshold Limit / Range
Ventilator TFT displays size (diagonal) in inch.	10 or more (higher size will be preferred)
Ventilator Tidal Volume in ml.	(20 -1400) or more
Ventilator frequency in breaths/minutes	(4 - 50) or more
Ventilator I:E ratio	(2:1 - 1:4) or more
Ventilator inspiratory pause in % of Time	0-60
Ventilator PEEP in cm H ₂ O	(0 / off -30) or more
Ventilator pressure limit in cm H ₂ O	(10-60) or more
Sodalime capacity in liters	1 or more
Drawers non lockable	2 or more
Provision for AGM (in-built or external)	Mandatory
Provision for vaporizers should be of same manufacturer	2
Touch screen, Electronic Gas Mixing	

Parameter / Feature	Threshold Limit / Range
MUST HAVE SPECIFICATION	
Parameter / Feature	Yes
Provision for delivery of oxygen, nitrous oxide and medical air [2 oxygen (1 Pipeline supply + 1 cylinder), 2 nitrous (1 Pipeline supply + 1 cylinder) and 1 air (1 Pipeline supply)]	
Circle absorber, ventilator and vaporizer (Universal Sodamine) FiO ₂ and ETO ₂ paramagnetic monitor	
Independent attachments and auto-switching facility for connecting central gas and pin indexed cylinders & DISS	
Non interchangeable pipe line hose	
Large size pressure gauge / Digital Display	
Integrated independent oxygen flow meter	
Quick availability of gases on the time of machine on/off	
Selector switch should be there for open and close circuit, bag to ventilator.	
The common gas outlet mounted at front of the machine	
Writing table/work surface	
Top shelf	
N ₂ O cut off facility if O ₂ supply fails	
O ₂ failure alarm both visual and audible	

Parameter / Feature	Threshold Limit / Range
Availability of O2 flush (20-75 Litres.) bypassing vaporiser	
Visual display of individual gas	
Maintenance free vaporiser for the life of the product	
Pneumatically / Electrically driven and electronically controlled Ventilator	
Control of ventilator parameter by touch screen and rotary dial	
Availability of VCV, PCV, SIMV, spontaneous/pressure support, manual modes. Peak Pressure and Plateau Pressure should be visible in display	
Flow triggered (0.3 – 10 liters) assist mode	
Leak compensation should be visible in the display (leak test should be user friendly)	
User adjustable alarms for major parameters	
User adjustable apnea alarms	
Display of patient loop (Pressure & Volume) and wave form (flow /volume and air way pressure), CO2 wave form.	
Display of compliance and resistance	
Separate flow sensor for inspiratory and expiratory breathing path	
Reusable & autoclavable type volume measurement sensor	

Parameter / Feature	Threshold Limit / Range
Latex free and autoclavable breathing system except O5 sensor with musk & HME filter : Adult: 2 nos., Paediatric: 2 nos. and Neonates 2 nos. & disposable 50 for each (Mask 0-5 all size – 2 each) (Breathing Bag (0.5 L-3 L) all size - 2 each)	
EtCO ₂ Sampling Line with T Connector for side stream: 50 nos.	
Acceding and clearly visible Bellows	
Provision for low flow anaesthesia	
Paramagnetic O ₂ sensor	
Anaesthesia Gas Monitoring Module ; USFDA / European CE (4 digit notified body) / BIS	
Electrical Safety and Quality Certification	
Standard & safety certification (European CE (4 digit notified body) / US FDA approved / BIS)	