



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

**Supply and Commissioning of Medical equipment for Mother and Child Hub at Canning Sub
Divisional Hospital, South 24 Pgs**
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-384/2022

Dated-29.08.2022

(2nd call to Bid Reference No.: WBMSCL /NIT-331/2022 Dated –25.07.2022)

The following amendment has been made in the bid document:

Amendment – I

(Page No.-34)

REVISED TECHNICAL SPECIFICATION

Item – 1

Cardiotocography (CTG) equipment

1. Cardio tocography equipment with auto-correlation.
2. CTG equipment should have Wave form display and digits simultaneously
3. Vibro acoustic stimulation and twin monitoring facilities desirable
4. CTG equipment should have Computer compatibility software facilities
5. CTG equipment should have Maturity sensed and automatic fetal movement detection.
6. CTG equipment should have Alert alarm for tachycardia and Bradycardia
7. CTG equipment should have Battery (Lithium / Alkaline) back up for at least 3 - 4 hrs.
8. Patients data storage for at least 10 hrs
9. CTG equipment Should have facility for 3 displays [2 display for FHR (TWIN) and 1 for Uterine pressure]
10. CTG equipment should be supplied with event marker facility for each foetal movement perceived & noted by the mother.

11. Should be supplied with a Trolley
12. Should be supplied with a Printer
13. Screen size within 8"- 12"
14. should be **European CE (4 digit) / USFDA / UKCA / UK Certification/ BIS** certified
15. Should supply 50 rolls each of CTG thermal paper (45-50 meter).
16. Conducting jelly / paste - 02 nos (200 ml each).
17. Continuous supply of paper roll
18. The rate of paper roll and other consumables should be quoted in Form 9 (a).