



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and commissioning of 6(six) of different medical equipment for Department of Urology at Calcutta National Medical College & Hospital
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-304/2025

Dated-17.04.2025

Amendment-II

Technical Specification

Schedule-I

Telescope (4 MM)

- a) Should be 4 mm diameter, Angle of view 30 degree
- b) Should have length of 30 cm
- c) Should be autoclavable
- d) Should be fiber optic light transmission incorporated
- e) Should have instrument tray for sterilization purposes with fixing system of telescope.
- f) Should have CDSCO approved.
- g) Should have 2 yrs manufacture warranty
- h) Should be compatible with existing resectoscope
- i) Should be compatible with existing working element