



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Procurement, supply and installation of different medical equipment in the Hospital and
Medical College of the Government of West Bengal
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-144/2025

Dated-20.02.2025

The following amendment have been made in the tender document,

Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Excimer 308 nm

1. Wavelength - 308 nm.
2. Spot Size - 40*40 mm/ 16 cm².
3. Medium Output Density - 50 mW/cm².
4. Energy Density - 50-4500 mJ/cm².
5. Power Requirements - 100-240 V, 0.8 A, 50/60 Hz.
6. **Certification:** Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable.