



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/319
E mail: procurement@wbmsc.gov.in / to@wbmsc.gov.in

Procurement of equipment /item for the Sub-Centre(SC), Primary Health Centre(PHC), Block Public Health Unit (BPHU) and Urban Primary Health Unit (UPHC) of the State of West Bengal under 15th Finance Commission.

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT- 144/2022

Dated-12.04.2022

The following amendment have been made in the tender document,

Amendment –I (Revision of Technical Specification)

The revised technical specifications for the item is given below,

Foetal Doppler

1. It should be LCD / **LED** Display (45mm X 22mm or more)
2. It should be ultrasonic emitting frequency: 2 – 2.5 MHz and power: < 5m Watt /cm²
3. Overall sensitivity at the distances 120 - 200 mm from the face of the probe: 90db
4. Automatic FHR test range 50 – 210 BPM or wider (Beat / Minute)
5. Resolution: 1 BPM
6. Audio output : 0.5 – 1 Watt
7. Should be supplied with electric charger & rechargeable battery (Li – ion) with 1 set of spare battery.
8. Good quality water proof case for carrying purpose

9. Good quality water proof probe & gel (5 bottles of 200ml or total quantity of 1000ml) should be supplied with the equipment.
10. Warranty includes the entire machine including probe.
11. **Standards, Safety & Training:**
 - a) Should be CE (“Conformite Europeene”) from European Union notified body having 4 digit identification number or US FDA certification. Only product enlistment under US FDA will not be accepted.
 - b) Warranty: 3 years.
 - c) On site comprehensive training for field staff and support services till customer satisfaction with the system.