

## Notice Inviting e-Tender

## West Bengal Medical Services Corporation Limited Swasthya Sathi GN-29, Salt Lake, Sector-V

Kolkata-700091

Phone No (033) 40340307/320 E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Different Medical Equipment for Calcutta National Medical College and Hospital

(Submission of Bid through online)

 Bid Reference No.: WBMSCL/NIT-210/2025
 Dated-18.03.2025

 2<sup>nd</sup> call of bid reference No. WBMSCL/NIT- 071/2025 Dated-21.01.2025
 Dated-18.03.2025

## Amendment-I (Schedule II)

## Schedule II Revised Technical Specification Automatic Slide Stainer

- 1. Instrument should have the following features:
- 2. High Throughput robotic liner stainer for multiple staining applications and should run up to 11 slide racks in parallel.
- 3. Specimen slide throughput: at least 200 specimen slides per hour (depending on the selected program- up to 600 slides per hour).
- 4. Simultaneous staining of various different staining protocols.
- 5. Exact incubation times with simple, menu-driven programming.
- 6. Incubation times and sequence of use of reagent stations must be freely programmable as per user's choice
- 7. Solvent resistant screen to monitor the staining process.
- 8. Racks should be assigned to the correct staining protocol.
- 9. At least total 26 stations with 18 reagent stations and maximum 5 wash stations of upto 450 ml capacity and racks with 20-30 specimen slides for high throughput.
- 10. Programmable for 15 programs of up to 25 steps each with incubation time setting from 0 sec to 59 min 59 sec / 99 min 59 sec
- 11. Integrated oven with temperature setting from 30 dg to 65 dg for optimal slide drying.
- 12. Continuous loading and unloading of slides via rack entry and exit door.

- 13. Specimen slide throughput of at least 200 slides per hour up to 600 slides per hour.
- 14. Agitation programmable from 0 to 20 times or continuous.
- 15. Regent containers can be exchanged quickly and easily.
- 16. Programmable up and down movement of robotic arm.
- 17. Integrated Fume extraction system having fan with charcoal filter to remove hazardous fumes.
- 18. Minimized user exposures to hazardous reagent fumes.
- 19. Gentle vibration to slide rack during lifting to reduce carry over contamination.
- 20. Easy-to-clean and resistant surfaces made out of polyester epoxy resin and stainless steel.
- 21. Audible warning buzzer in case of any error during operation.
- 22. Should be an UPGRADABLE Model: Must have the provision to be able to get attached to Automated Robotic cover slipper by just removing one side wall of the Linear Stainer/or must provide adaptor for the side rack to make it compatible with any make of automatic cover slipper, making it a WORKSTATION PLATFORM.
- 23. The bidder should submit valid CDSCO Certificate / Registration / License for both the manufacturer(s) and importer(s) as applicable.
- 24. At least 25 nos. of installations of automatic slide stainer in India out of which 15 nos. should be installed in Govt. Institution (Central and State) rest 10 nos. Should be installed in Govt. or Private institution.
- 25. The bidder should submit 10 nos. of the Satisfactory Service and Performance Report from only the Govt. Institutions otherwise the bid will not be treated as valid.
- 26. This service support should be available within 48 hours from the time of logging information and the machine should be fully functional mode.