



## Notice Inviting e-Tender

**West Bengal Medical Services Corporation Limited**  
**Swasthya Sathi**  
**GN-29, Salt Lake, Sector-V**  
**Kolkata-700091**

Phone No (033) 40340307/320  
E mail: procurement@wbmsc.gov.in

**Procurement of Medical Items for the Blood Centers at Hospitals and Medical Colleges of Govt. of West Bengal**  
(Submission of Bid through *online*)

<b>Bid Reference No.: WBMSCL/NIT-175/2025</b>	<b>Dated-01.03.2025</b>
<b>2<sup>nd</sup> call of bid reference no. WBMSCL/NIT-193/2024 Dated-11.03.2024</b>	

### Amendment-I

## Revised Technical Specification

### Item-3

#### Table Top Tube Sealer

Feature	Specification
Power supply input frequency	100V AC to 240V AC 50Hz/60Hz
Classification	Protection against electric shock
Operation mode	Once it is turned on , the device is ready to use
RF output power (effective)	20W
Indicator lights / display	Cover open(Yellow), Power(green), Ready(green), Seal(red) or LCD display
Tube detection	Automatic
RF output frequency	40.68 Mhz
Tube specification	Polyvinyl chloride (PVC) tubing up to 6mm (outer diameter)
Sealing time	1 to 3 seconds, adjustable
Dimensions(W*d*h)	155mm x 330mm x 160mm
Weight	5 to 6 kg
Power consumption	Operationg:185W Standby:20W
Fuse	250V,3.15A,type T, slow blow
Warranty	2 years

#### Certifications:

Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable.

**Note:**

1. Installation Quality (IQ), Operational Quality (OQ), Performance Quality (PQ) certificate to be submitted during installation of the machine.
2. 3<sup>rd</sup> party calibration certification should be also be provided during installation