



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and Commissioning of Medical equipment for the Department of Urology at R.G.Kar MCH
(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-206/2022

Dated-19.05.2022

Amendment -II

REVISED TECHNICAL SPECIFICATION

Schedule – IV

Semi Rigid Uretero Renoscope with all accessories

1. 6.0-8.0 Fr Uretero-roscope – 02 (Two)

2. Alligator Forceps – 03 (three)
3. Tri-pronged forceps – 03 (three)

Specification

- The Semi-rigid Uretero-roscope should have a direction of view between 5 to 7 degrees.
- The distal end outer diameter of the Uretero-roscope should be between 6 to 8 Fr.
- The working length of the Uretero-roscope should be around 400-450 mm.
- The diameter of the forceps channel should be around 4-5 Fr.
- The Uretero-roscope should be a Semi-Rigid scope.
- The tip of the Uretero-roscope should be atraumatic.
- The Uretero-roscope should be autoclavable.
- The Uretero-roscope should have built-in maintenance free stop cocks.

- The Uretero-roscope should be supplied with the following items:
 - a) Durable detachable bridge for irrigation.
 - b) 10 self-sealing caps.
 - c) Light guide adapters.
 - d) Along with all accessories. i.e. rigid & flexible forceps, tri-prong forceps, biopsy forceps, pathfinder, brush for cytology, disinfectant tray. (Each items of 2 in nos.)
 - e) All the accessories should be compatible with the quoted uretero-roscope
 - f) The light source head attachment should be such that it can be attached to any light source with or without adaptor.
 - g) It should be from the same OEM.
- Should be **European CE (4 Digit Notified body) & USFDA** certified.