



Notice Inviting e-Tender

West Bengal Medical Services Corporation Limited
Swasthya Sathi
GN-29, Salt Lake, Sector-V
Kolkata-700091

Phone No (033) 40340307/320
E mail: procurement@wbmsc.gov.in

Supply and commissioning of various ophthalmic equipment across different Health Facilities of the Govt of West Bengal under the project NPCB and VI

(Submission of Bid through *online*)

Bid Reference No.: WBMSCL/NIT-434/2025

Dated-24.05.2025

2nd call for items at Schedule No .V, VI, VII, VIII, IX, XI,XII and XIII of
Bid Reference Number WBMSCL/NIT-293/2025 Date-16.04.2025

Amendment I

Revised Technical Specification

Schedule No.V Gonioscope

1. It should be 3 mirror lens.
2. It should be two opposing 62 degree mirrors providing a complete view of the anterior chamber angle with only a 180 degree lens rotation.
3. It should be available for use without methyl cellulose (NMR).
4. Lens magnification should be at least 0.80x.
5. Contact diameter should be 15 mm.
6. Lens height should be at least 20 mm.

Safety & Standard: Valid CDSCO Certificate License for the manufacturer(s) & Registration Certificate from CDSCO for the importer(s) as applicable